



MEDISCA[®]
YOUR TRUSTED PARTNER IN COMPOUNDING

PRODUCT NO.: 5766



Hot Plate/Stirrer, 10 x 10" (25.4 x 25.4cm) (Guardian 5000) (230V)

DESCRIPTION: The Ohaus Guardian 5000 Hotplate Stirrers are designed for safety and simplicity for everyday use. The design ensures proper mixing and temperature control of the top plate or your sample when using the optional temperature probe.

FEATURES:

- ✔ An industry-leading early detection system that uses two independent safety controls to continuously monitor the electronics and shut off heating before an over temperature condition occurs
- ✔ Precise stirring from 60 to 1600 RPM is made possible by the design of a powerful motor, strong magnet and software controlled ramp rate, ensuring secure magnetic coupling for viscous applications.
- ✔ The bright LCD display includes large temperature and speed settings, intuitive icons and a prominent hot top indicator - all designed to quickly and easily monitor performance from across the lab.
- ✔ The SmartHousing is chemical resistant with an innovative, easy-to-clean design that channels spills away from internal components and the angled control panel.

TECHNICAL SPECIFICATIONS:

Capacity	18 L
Plate Construction	Ceramic
Plate Size	10 in x 10 in (254mm x 254 mm)
Temperature Range	Ambient +5°C – 500°C
Control	Digital
Display	Customer Backlit LCD
Weight	11.5 lb (5.2 kg)
Power	230V
Speed Range	60 to 1600 RPM
Dimensions, L x W x H (mm/in)	415 x 286 x 122/16.3 x 11.25 x 4.8

USA | CANADA | AUSTRALIA PACIFIC | EUROPE

WWW.MEDISCA.COM

WARRANTY: 2 years

USA | CANADA | AUSTRALIA PACIFIC | EUROPE

WWW.MEDISCA.COM

Please note that the information contained herein is for informational purposes only. The information is based on recognized compendia, manufacturers' instructions and/or other authoritative sources and texts. MEDISCA takes no responsibility for the validity or accuracy of this information. Customer MUST refer to USP, FDA, Health Canada, TGA and/or their State Board to ensure they comply with the specific regulations and ensure that appropriate procedures are in place.